

Fox Valley Metrology

3114 Medalist Drive

Oshkosh, WI 54902

(920) 426-5894 • Fax (920) 426-8120

<http://www.FoxValleyMetrology.com>**CERTIFICATE OF
CALIBRATION**

Certificate No. ACT-1272

CERTIFICATION NUMBER CL011-37670-434**FOR** Intertek
8431 Murphy Drive
Middleton, WI 53562**PURCHASE ORDER #****TEST INSTRUMENT** Differential Pressure Transducer**MAKE** Omega**MODEL** PX263-01D5V (Intertek)**SERIAL NUMBER** X15160229**IDENTIFICATION** 1406 **CUSTOMER LOCATION****CONDITION RECEIVED** In Tolerance**CONDITION RETURNED** In Tolerance**CALIBRATED BY** Bob Mooney**FORM REVIEWED BY** M.T.**CALIBRATION LOCATION** On Site**ENVIRONMENT** 70.0°F, 21.1°C, 21.0%RH**CALIBRATION DATE** 01/11/2021**RECALIBRATION DUE** 07/11/2021**PROCEDURES FOLLOWED**

FVE-060 rev. 0

This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated.

Form Revision 7: 07/14/2020

STANDARDS USED

INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL
FVS-708	N/A	CK272-27052-397	09/30/2021
FVS-789	7543203	CK014-22597-466	02/28/2021
FVS-789A	7461971	CK014-22979-466	02/28/2021

Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.

All instruments have been calibrated against standards traceable to NIST. Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
Pressure/Vacuum	(inH20)	(inH20)	(inH20)	(inH20)	(inH20)	(inH20)
Low	-0.0000	-0.0025	0.0025	0.0000	0.0000	0.00006
Vacuum	-0.0500	-0.0525	-0.0475	-0.0503	-0.0503	0.00006
	-0.1000	-0.1025	-0.0975	-0.0998	-0.0998	0.00006
	-0.5000	-0.5025	-0.4975	-0.5002	-0.5002	0.00007
	-1.0000	-1.0025	-0.9975	-0.9982	-0.9982	0.00006
High	0.0000	-0.0025	0.0025	0.0000	0.0000	0.00006
pressure	0.0500	0.0475	0.0525	0.0504	0.0504	0.00007
	0.1000	0.0975	0.1025	0.1005	0.1005	0.00007
	0.5000	0.4975	0.5025	0.5010	0.5010	0.00007
	1.0000	0.9975	1.0025	1.0007	1.0007	0.00007